

IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-M2-P
S.N.: 227
Date: Aug 13 79

- a) Output signal amplitude:
 > 60 VOLTS
b) Pulse width:
 2.0 nSEC
 (10% RISE)
c) Jitter, stability:
 OK
d) Prime power:

+135V
5 mA

[Handwritten signature]

Aug
D.S.

2.0 nSEC/DIV
≈ 16.0 VOLTS/DIV