



AVR-E4-B-P, at maximum & minimum pulse widths.  
100 V/div, 1 ns/div, 100 kHz.

- 20, 50 or 100 Volt peak outputs
- sub-nanosecond rise times
- 1 ns to 5 us pulse widths
- PRF to 1 MHz
- IEEE-488.2 GPIB control
- Ethernet optional

The AVR-E family offers medium-to-high voltage pulses with very fast rise and fall times.

The AVR-E1 series provides peak outputs to 20V, with pulse widths variable from 10 to 200 ns, pulse repetition frequencies to 200 kHz, and rise times of 300 ps. The minimum pulse width can optionally be reduced to 1 ns. The maximum pulse width can optionally be increased to 5 us, with a reduction in the maximum PRF to 20 kHz.

The AVR-E2 series has similar pulse width ratings, but offers a higher maximum amplitude of 50V, with 500 ps rise times and 1 ns fall times.

The AVR-E3 series offers a maximum amplitude of 100V, with 500 ps rise time, 1 ns fall time, and a maximum PRF of 100 kHz (or 20 kHz with the wider pulse width options).

The AVR-E3A series also provides a maximum amplitude of 100 V but with a 1.2 ns rise, a 10 to 500 ns pulse width range and a maximum PRF of 200 kHz.

The narrow-pulse AVR-E4 series can generate 100V pulses with pulse widths variable from 1 to 5 ns, at repetition rates to 100 kHz. The AVR-E5 series covers the pulse width range of 1 to 10 ns, with 50V maximum amplitude and a maximum repetition rate of 1 MHz.

The AVR-E6 series covers the pulse width range of 8 to 30 ns, with 100V maximum amplitude and a maximum repetition rate of 2 MHz.

All models include a complete computer control interface. This provides GPIB and RS-232 computer-control, as well as front panel keypad and adjust knob control of the output pulse parameters. (For additional details, please see <http://www.avtechpulse.com/gpib>.) A large backlit LCD displays the output amplitude, frequency, pulse width, and delay. To allow easy integration into automated test systems, the programming command set is based on the SCPI

standard. LabView drivers are available for download at <http://www.avtechpulse.com/labview>.

The -VXI option adds a rear-panel Ethernet connector, allowing the instrument to be remotely controlled using the VXI-11.3, ssh, telnet, and web protocols. In particular, the VXI-11.3 features allows software like LabView to control an instrument using standard VISA communications drivers and network cabling, instead of using older-style GPIB cabling and GPIB controller cards. For additional details, please see <http://www.avtechpulse.com/options/vxi>.

All models are protected from overload conditions (such as excessively high duty cycle or short circuited load) by an automatic control feature that limits the output power for as long as the overload condition exists.

A delay control and a sync output are provided for scope triggering purposes. The units can also be triggered externally using a TTL-level pulse. Either output polarity or an optional dual output polarity can be provided.

A DC offset or bias insertion option is available. Units with this option include a circuit similar to Model AVX-TC (see <http://www.avtechpulse.com/bias/avx-tc> for details) at the output. The required DC offset or bias is applied directly to rear panel solder terminals.

Visit <http://www.avtechpulse.com> for application notes, data sheets, LabView drivers, pricing, and more!

Many models can be customized to meet specific requirements. Contact Avtech engineers at [info@avtechpulse.com](mailto:info@avtechpulse.com) for details.

Typical waveforms from actual production units are available online. See the individual product pages at:

<http://www.avtechpulse.com/speed>  
<http://www.avtechpulse.com/medium>



## SPECIFICATIONS

## AVR-E SERIES

Model <sup>1</sup> :	AVR-E1-B	AVR-E2-B	AVR-E3-B	AVR-E3A-B	AVR-E4-B	AVR-E5-B	AVR-E6-B
Maximum amplitude <sup>2</sup> : (50Ω load <sup>7</sup> )	20V	50V	100V	100V	100V	50V	100V
Rise time (20%-80%):	≤ 0.3 ns	≤ 0.5 ns	≤ 0.5 ns	≤ 1.2 ns	≤ 0.4 ns	≤ 0.5 ns	≤ 2.0 ns
Fall time (80%-20%):	≤ 0.6 ns	≤ 1.0 ns	≤ 1.0 ns	≤ 2.0 ns <sup>6</sup>	≤ 0.6 ns	≤ 1.0 ns	≤ 2.0 ns
Pulse width (FWHM):	standard units: 10 ns to 200 ns with -W1 option: 1 ns to 200 ns with -W2 option: 50 ns to 5 us with -W3 option: 1 ns to 5 us			10 to 500 ns	1 to 5 ns	1 to 10 ns	8 to 30 ns
Maximum PRF: Standard units, or with -W1 option: Units with -W2 or -W3 options:	200 kHz 20 kHz		100 kHz 20 kHz	200 kHz	100 kHz	1 MHz	2 MHz
Maximum duty cycle:	5%			2%	N/A		
Polarity:	Positive or negative or both (specify <sup>3</sup> )						
DC offset or bias insertion:	Option available <sup>4</sup> . Apply required DC offset or bias in the range of ± 25 Volts, (250 mA max) to back panel solder terminals.						
Trigger modes:	Internal trigger, external trigger (TTL level pulse, > 10 ns, 1 kΩ input impedance), front-panel "Single Pulse" pushbutton, or single pulse trigger via computer command.						
Variable delay (Sync to main out):	0 to 1.0 seconds, for all trigger modes (including external trigger).						
Propagation delay:	≤ 160 ns (Ext trig in to pulse out)						
Jitter: (Ext trig in to pulse out)	± 35ps ± 0.015% of sync delay						
Sync output:	> +3 Volts, > 50 ns, will drive 50 Ohm loads						
Gate input:	Synchronous or asynchronous, active high or low, switchable. Suppresses triggering when active.						
Connectors:	Out: SMA, Other: BNC						
Power requirements:	100 - 240 Volts, 50 - 60 Hz						
GPIB and RS-232 control <sup>1</sup> :	Standard on -B units.						
LabView drivers:	Check <a href="http://www.avtechpulse.com/labview">http://www.avtechpulse.com/labview</a> for availability and downloads						
Ethernet port, for remote control using VXI-11.3, ssh, telnet, & web:	Optional <sup>5</sup> . Recommended as a modern alternative to GPIB / RS-232. See <a href="http://www.avtechpulse.com/options/vxi">http://www.avtechpulse.com/options/vxi</a> for details.						
Settings resolution:	The resolution of the timing parameters (pulse width, delay, period) varies, but is always better than 0.15% of ( set value  + 20 ns). The amplitude resolution is < 0.1% of the maximum amplitude.						
Settings accuracy, after 10 minute warm-up:	Amplitude: Typically ± (3% of setting) ± (2% of maximum). Delay, Period: Typically ± (3% of setting) ± (5 ns) Pulse width: Typically ± (3% of setting) ± (2 ns), at maximum amplitude. As the amplitude is reduced, the pulse width may shift ± 5 ns. For high-accuracy applications requiring traceable calibration, verify the output with a calibrated oscilloscope <sup>8</sup> .						
Dimensions (H x W x D):	100 mm x 430 mm x 375 mm (3.9" x 17" x 14.8")						
Chassis material:	cast aluminum frame and handles, blue vinyl on aluminum cover plates						
Temperature range:	+5°C to +40°C						

- 1) -B suffix indicates IEEE-488.2 GPIB and RS-232 control of amplitude, pulse width, PRF and delay (see <http://www.avtechpulse.com/gpib>).
- 2) For operation at amplitudes of less than 10% of full-scale, best results will be obtained by setting the amplitude near full-scale and using external attenuators on the output.
- 3) Indicate desired polarity by suffixing model number with -P or -N (i.e. positive or negative) or -PN for dual polarity option.
- 4) The -OS option adds an internal bias tee circuit to the output, allowing an externally-generated DC offset (±25V/250mA max) to be added the output signal. The -OT option includes the -OS function, and adds the ability to generate a ±5V/100mA offset internally (controlled from the front panel, or by computer command).
- 5) Add the suffix -VXI to the model number to specify the Ethernet port..

- 6) Fall time increases to < 3 ns for pulse widths less than 15 ns.
- 7) A 50Ω load is required. Other loads may damage the instrument. Consult Avtech ([info@avtechpulse.com](mailto:info@avtechpulse.com)) if you need to drive other load impedances.
- 8) These instruments are provided with a basic calibration checksheet, showing a selection of measured output parameters. These measurements are performed with equipment that is calibrated on a regular basis by a third-party ISO/IEC 17025:2005 accredited calibration laboratory. However, Avtech itself does not claim any accreditation. For applications requiring traceable performance, use a calibrated measurement system rather than relying on the accuracy of the pulse generator settings.



AVR-E3-B