


IMPULSE GENERATOR PERFORMANCE CHECK


Model: AVH-5

S.N.: 105

Date: 26/11/76



1 MHz (HP 1817A SCOPE)



3 MHz (HP 1817A SCOPE)

- a) Output signal amplitude > 30 volts
- b) Pulse width ≤ 1 msec. (10% amplitude)
- c) Prime power 120 mA, 15 V.
- d) Jitter, stability OK

[Handwritten signature]