


IMPULSE GENERATOR PERFORMANCE
CHECK

Model: AVH-MP1

S.N.: 106

Date: DEC 20 '76



1 μ sec/DIV, 6.6 volts/DIV

- a) Output signal amplitude 33 volts
- b) Pulse width 0.82 μ sec
(10% R.T.)
- c) Prime power 15 volts, 75 mA
- d) Jitter, stability OK