


IMPULSE GENERATOR PERFORMANCE  
CHECK

Model: AVH-M1

S.N.: 115

Date: 19/3/77



1.0 nsec/div  
7 VOLTS/DIV.

- a) Output signal amplitude: 35 VOLTS.
- b) Pulse width:  $\leq 1.0$  nsec  
(10% RISE TIME)
- c) Prime power: +135 V  
4.0 mA
- d) Jitter, stability: OK.