


PULSE GENERATOR PERFORMANCE
CHECK


Model: *AVI-V*

S.N.: *117*

Date: *6/7/77*



12 VOLTS/DIV, 20 NSEC/DIV



12 VOLTS/DIV, 1 NSEC/DIV

- a) Output signal amplitude:
51 volts.
- b) Pulse width: *0-100 nsec*
(BACK PULSED FOR 90-100 NSEC)
- c) Rise time: *≤ 0.5 nsec.*
- d) Fall time: *≈ 1.0 nsec.*
- e) PRF: *20 KHZ.*
- f) Jitter, stability:
OK
- g) Prime power:
+15 volts
225 mA.