


IMPULSE GENERATOR PERFORMANCE  
CHECK

Model: AVH-M1  
S.N.: 120  
Date: 20/6/77



- a) Output signal amplitude:  $\approx 31$  VOLTS
- b) Pulse width:  $\leq 1.0$  NSEC.  
(10% RISE TIME)
- c) Prime power  $\pm 135$  V  
 $2.8$  mA
- d) Jitter, stability:

OK

1.0 NSEC/DIV.  
7 VOLTS/DIV.