



PULSE GENERATOR PERFORMANCE
CHECK

Model: *AUI-V*
S.N.: *121*
Date: *19/7/77*



11 volts/div, 20 nsec/div



100 nsec/div

- a) Output signal amplitude: *52 volts.*
- b) Pulse width: *0-100 nsec.*
- c) Rise time: ≤ 0.5 nsec
- d) Fall time: ≤ 2.0 nsec
- e) PRF: *20 kHz*
- f) Jitter, stability: *OK.*
- g) Prime power: *+15 volts.*
240 mA

[Handwritten signature]