


PULSE GENERATOR PERFORMANCE  
CHECK

Model: *AVI-V*  
S.N.: *122*  
Date: *2/9/77*



*12 VOLTS / DIV.*  
*10 NSEC / DIV.*

- a) Output signal amplitude: *52 VOLTS.*
- b) Pulse width: *0-100 nsec*
- c) Rise time: *0.5 nsec.*
- d) Fall time: *2.0 nsec.*
- e) PRF: *≤ 20 KHZ*
- f) Jitter, stability: *OK*
- g) Prime power: *+15 ± 1 VOLTS*  
*< 250 mA*