


PULSE GENERATOR PERFORMANCE
CHECK

Model: AVI-V

S.N.: 123

Date: 2/9/77



- 12 VOLTS/DIV.
10 NSEC/DIV.

- a) Output signal amplitude: 50 VOLTS
- b) Pulse width: 6 - 100 NSEC
- c) Rise time: 0.5 NSEC
- d) Fall time: ≤ 2 NSEC
- e) PRF: ≤ 20 KHZ
- f) Jitter, stability: OK
- g) Prime power: $+15 \pm 1$ VOLT
 < 250 mA