


PULSE GENERATOR PERFORMANCE
CHECK

Model: AVI-V

S.N.: 124

Date: 19/9/77



12 VOLTS/CM.
20 NSEC/DIV.

- a) Output signal amplitude: 51 VOLTS
- b) Pulse width: 0-100 NSEC
- c) Rise time: 0.5 NSEC
- d) Fall time: ≤ 2.0 NSEC
- e) PRF: ≤ 20 KHZ
- f) Jitter, stability: OK.
- g) Prime power $+15 \pm 1$ VOLT
 < 250 mA