


PULSE GENERATOR PERFORMANCE
CHECK

Model: *AVI-V*

S.N.: *125*

Date: *19/9/77*



12 VOLTS/cm
20 nsec/cm

- a) Output signal amplitude: *51 VOLTS*
- b) Pulse width: *0-100 nsec.*
- c) Rise time: *0.5 nsec*
- d) Fall time: *≤ 2.0 nsec.*
- e) PRF: *0-20 KHz.*
- f) Jitter, stability: *OK*
- g) Prime power *+15 ± 1 VOLTS.*
≤ 250 mA