


IMPULSE GENERATOR PERFORMANCE
CHECK

Model: *AVH-MIP*

S.N.: *131*

Date: *18/11/77*



a) Output signal amplitude

> 30 VOLTS.

b) Pulse width

1.0 NSEC (10% RISE TIME)

c) Prime power

*+15 ± 1 VOLT
135 mA*

d) Jitter, stability

OK -