


PULSE GENERATOR

PERFORMANCE      CHECK

Model: *AVI*

S.N.: *132*

Date: *18/11/77*



*2 NSEC / DIV.*  
*20 VOLTS / DIV.*

- a) Output signal amplitude: *> 40 VOLTS.*
- b) Pulse width: *VARIABLE FROM 1.0 TO 100 NSEC.*
- c) Rise time: *≤ 0.5 NSEC.*
- d) Fall time: *≤ 2.0 NSEC.*
- e) PRF: *OK.*
- f) Jitter, stability: *OK.*
- g) Prime power: *+15 ± 1 VOLT.*  
*⊙ 135 mA.*