


IMPULSE GENERATOR PERFORMANCE

CHECK

Model: AVH-MP
S.N.: 135
Date: 16/12/77



- a) Output signal amplitude
 $\gg 30$ VOLTS.
- b) Pulse width ≤ 1.0 NSEC
 (< 10% RISE TIME)
- c) Prime power $\pm 15 \pm 1$ VOLT
 135 mA
- d) Jitter, stability

E.R.

! NSEC / CM?