


PULSE GENERATOR PERFORMANCE

CHECK

Model: AVI-V
S.N.: 136
Date: 19/1/78



10 NSEC/DIV.

- a) Output signal amplitude ≥ 50 VOLTS
- b) Pulse width 0-100 NSEC
- c) Rise time 0.5 NSEC
- d) Fall time ≤ 2.0 NSEC
- e) PRF ≤ 20 KHZ
- f) Jitter, stability OK
- g) Prime power $+15 \pm 1$ VOLT
 ≤ 200 mA