


PULSE GENERATOR PERFORMANCE

CHECK

Model: AVI-V
S.N.: 137
Date: 19/1/78



10 nsec/div

- a) Output signal amplitude
 ≥ 50 VOLTS
- b) Pulse width
0-100 nsec
- c) Rise time
 ≤ 0.5 nsec
- d) Fall time
 ≤ 2.0 nsec
- e) PRF
 ≤ 20 KHz
- f) Jitter, stability
OK.
- g) Prime power
 $+15 \pm 1$ VOLT
 ≤ 200 mA