


PULSE GENERATOR PERFORMANCE

CHECK

Model: AVI-V

S.N.: 140

Date: 23/1/78



10 NSEC/DIV

- a) Output signal amplitude  
     $\geq 50$  VOLTS
- b) Pulse width  
    0 - 100 NSEC
- c) Rise time  
     $\leq 0.5$  NSEC.
- d) Fall time  
     $\leq 2.0$  NSEC.
- e) PRF  
     $\leq 20$  KHz.
- f) Jitter, stability  
    OK
- g) Prime power  
     $+15 \pm 1$  VOLT  
     $\leq 200$  mA