


PULSE GENERATOR PERFORMANCE

CHECK

Model: AVI-V  
S.N.: 141  
Date: 23/1/78



- a) Output signal amplitude  
 $\geq 50$  VOLTS
- b) Pulse width  
0-100 NSEC.
- c) Rise time  
 $\leq 0.5$  NSEC.
- d) Fall time  
 $\leq 2.0$  NSEC
- e) PPF  
 $\leq 20$  KHZ
- f) Jitter, stability  
OK
- g) Prime power  
 $+15 \pm 1$  VOLT  
 $\leq 200$  mA