


PULSE GENERATOR PERFORMANCE
CHECK

Model: *AVI-V*

S.N.: *146*

Date: *17/4/78*



20 NSEC / DIV.
≈ 13 VOLTS / DIV.

a) Output signal amplitude:

≥ 50 VOLTS

b) Pulse width:

0-100 NSEC

c) Rise time:

≤ 0.5 NSEC.

d) Fall time:

≤ 2.0 NSEC

e) PRF:

≤ 20 KHz

f) Jitter, stability:

OK.

g) Prime power:

± 15 ± 1 VOLT

≤ 200 mA