


IMPULSE GENERATOR PERFORMANCE

CHECK

Model: AVH
S.N.: 148
Date: 1/5/78



1.0 NSEC/DIV
≈ 7 VOLTS/DIV.

- a) Output signal amplitude
 ≈ 30 VOLTS
- b) Pulse width ≤ 1.0 NSEC
 (10% RISE TIME)
- c) Prime power $+15 \pm 1$ VOLT
 135 mA
- d) Jitter, stability

OK