



PULSE GENERATOR PERFORMANCE

CHECK

Model: AVI-V
S.N.: 153
Date: 15/5/78



10 nSEC/DIV
≈ 15 VOLTS/DIV



10 nSEC/DIV
≈ 13 VOLTS/DIV

- a) Output signal amplitude
 ≥ 50 VOLTS
- b) Pulse width
 0-100 nSEC
- c) Rise time
 ≤ 0.5 nSEC
- d) Fall time
 ≤ 2.0 nSEC
- e) PRF
 ≤ 20 KHz
- f) Jitter, stability
 OK
- g) Prime power
 $+15 \pm 1$ VOLT
 ≤ 200 mA

A handwritten signature in black ink, consisting of several loops and a long horizontal stroke.