


IMPULSE GENERATOR PERFORMANCE

CHECK

Model: AVH-MIP

S.N.: 158

Date: 15/8/78



1 nsec/div  
x 6 VOLTS/DIV

- a) Output signal amplitude  
 $\geq 30 \text{ V}$
- b) Pulse width  $\leq 1.0 \text{ nsec}$   
(10% RISE TIME)
- c) Prime power  
 $\pm 15 \pm 1 \text{ V.}$
- d) Jitter, stability  
OK.

*[Handwritten signature]*