


PULSE GENERATOR PERFORMANCE

CHECK


Model: AVK-AV-1P

S.N.: 154

Date: 26/7/78



MAX OUTPUT
1.0 NSEC / CM



MID RANGE OUTPUT

- a) Output signal amplitude:
 ≤ 15 VOLTS
- b) Pulse width:
 0 - 50 NSEC.
- c) Rise time:
 ≤ 150 NSEC
- d) Fall time:
 ≤ 1 NSEC
- e) PRF:
 50 KHZ
- f) Jitter, stability:
 OK
- g) Prime power:
 $+15 \pm 1$ VOLT
 < 300 mA.