


PULSE GENERATOR PERFORMANCE

CHECK


Model: AVK-A-P

S.N.: 163

Date: OCT 12 78



7.0 V/CM
1.0 nsec/CM



a) Output signal amplitude:

≈ 15 VOLTS

b) Pulse width:

4 - 100 nsec.

c) Rise time:

≤ 15 nsec.

d) Fall time:

OK

e) PRF:

OK

f) Jitter, stability:

OK

g) Prime power:

$\pm 15 \pm 1$ VOLTS

250 mA.

A handwritten signature or set of initials, possibly "J.C.", written in dark ink.