


PULSE GENERATOR PERFORMANCE

CHECK


Model: AVK-AV-N

S.N.: 164

Date: OCT 23 78



MID RANGE OUTPUT
1.0 nsec/cm



MAX OUTPUT
1.0 nsec/cm

- a) Output signal amplitude
 > 15 VOLTS.
- b) Pulse width
 0 - 50 nsec.
- c) Rise time
 ≤ 150 psec.
- d) Fall time
- e) PRF: OK
- OK.
- f) Jitter, stability: OK
- g) Prime power:
 +15 ± 1 VOLT
 < 300 mA

[Handwritten Signature]