


PULSE GENERATOR PERFORMANCE

CHECK

Model: AVH-5

S.N.: 166

Date: NOV 3 78



a) Output signal amplitude  
 $\geq 15$  VOLTS

b) Pulse width:  
 $\leq 0.5$  NSEC

c) Rise time:  
OK

d) Fall time:  
OK

e) PRF:  
 $\leq 1.0$  MHz

f) Jitter, stability:  
OK

g) Prime power  
 $+ 15 \pm 1$  VOLT  
160 mA

1 NSEC/DIV  
 $\approx 10$  VOLTS/DIV