


PULSE GENERATOR PERFORMANCE

CHECK


Model: AVK-V

S.N.: 168

Date: NOV 16 78



1.0 nsec/div
6.0 volts/div



10.0 nsec/div
6.0 volts/div

- a) Output signal amplitude
 $> 15 \text{ V}$
- b) Pulse width:
 $0 - 50 \text{ nsec}$
- c) Rise time:
 $\leq 150 \text{ psec}$
- d) Fall time:
OK
- e) PRF:
 $0 - 50 \text{ kHz}$
- f) Jitter, stability:
OK
- g) Prime power $+15 \text{ V} \pm 1 \text{ volt}$
 275 mA

A handwritten signature in black ink, consisting of several loops and strokes.