


PULSE GENERATOR PERFORMANCE

CHECK


Model: *AVK-V*

S.N.: *169*

Date: *NOV 16 1978*



*1.0 nsec/div  
≈ 6.0 volt/div*



*10.0 nsec/div  
≈ 6.0 volt/div*

- a) Output signal amplitude  
*> 20 VOLTS*
- b) Pulse width:  
*0 - 50 nsec*
- c) Rise time:  
*≤ 150 nsec*
- d) Fall time:  
*OK*
- e) PRF:  
*0 - 50 kHz*
- f) Jitter, stability:  
*OK*
- g) Prime power  
*+15 ± 1 VOLT  
310 mA*

*[Signature]*