

one copy


PULSE GENERATOR PERFORMANCE

CHECK

Model: *AVI-V*

S.N.: *177*

Date: *29 DEC 78*



*10 NSEC/DIV*

*~ 15 VOLTS/DIV*

a) Output signal amplitude:

*50 VOLTS*

b) Pulse width:

*0 - 100 NSEC*

c) Rise time:

*≤ 0.5 NSEC*

d) Fall time:

*≤ 2.0 NSEC*

e) PRF:

*0 - 20 KHZ*

f) Jitter, stability:

*OK*

g) Prime power:

*+ 15 ± 1 VOLT*

*250 mA*