


PULSE GENERATOR PERFORMANCE

CHECK

Model: *AVI-V*

S.N.: *182*

Date: *FEB 8 79*



*10 nsec / DIV*  
*≈ 15 VOLTS / DIV*

- a) Output signal amplitude:  
*50 VOLTS.*
- b) Pulse width:  
*0-100 nsec.*
- c) Rise time:  
*≤ 0.5 nsec.*
- d) Fall time:  
*≤ 2.0 nsec*
- e) PRF:  
*0-20 KHz*
- f) Jitter, stability:  
*OK.*
- g) Prime power:  
*+15 ± 1 VOLTS @*  
*210 mA*

*[Handwritten Signature]*