


PULSE GENERATOR PERFORMANCE

CHECK

Model: *AVI-MP*
S.N.: *191*
Date: *MAR 22 79*



1.0 nsec/div

- a) Output signal amplitude
7.40 VOLTS
- b) Pulse width:
0.8 to 100 nsec
- c) Rise time:
≤ 500 psec.
- d) Fall time:
≤ 2.0 nsec
- e) PRF:
OK.
- f) Jitter, stability:
OK.
- g) Prime power:
+15 ± 1 VOLT
150 mA