


PULSE GENERATOR PERFORMANCE

CHECK

Model: AVL-AU  
S.N.: 194  
Date: APRIL 6 79



- a) Output signal amplitude:  
50 - 250 VOLTS
- b) Pulse width:  
0 - 100 NSEC
- c) Rise time:  
≤ 5 NSEC
- d) Fall time:  
≤ 5 NSEC
- e) PRF:  
0 - 5 KHZ
- f) Jitter, stability:  
OK.
- g) Prime power:

+15 VOLTS  
(±10)  
500 mA Ⓢ  
5 KHZ.

≈ 80 VOLTS/DIV  
10 NSEC/DIV.

↑  
used in manual  
later sheet