


PULSE GENERATOR PERFORMANCE

CHECK

Model: *AVI-V*
S.N.: *195*
Date: *APRIL 6 1979*



10 nsec / DIV
≈ 17 volts / DIV

- a) Output signal amplitude:
> 50 VOLTS
- b) Pulse width:
0 - 100 nsec
- c) Rise time:
≤ 0.5 nsec
- d) Fall time:
≤ 2.0 nsec.
- e) PRF:
0 - 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 ± 1 VOLT
250 mA