


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-S

S.N.: 196

Date: APRIL 16 1979



a) Output signal amplitude:

$> 15$  VOLTS

b) Pulse width:

$\leq 0.5$  nsec  
(10% RISE TIME)

c) Jitter, stability:

OK

d) Prime power:

$+15 \pm 1$  VOLT.

165 mA

1.0 nsec/DIV.

$\times 7.0$  VOLTS/DIV.