


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-M2

S.N.; 201

Date: APRIL 21 1979



- a) Output signal amplitude:
 > 60 VOLTS
- b) Pulse width:
 < 2.0 NSEC
 (10% RISE TIME)
- c) Jitter, stability:
 OK.
- d) Prime power:
 +135 VOLTS.
 +5 mA

2.0 NSEC/DIV
≈ 16 VOLTS/DIV.