


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVK-A-P*

S.N.: *202*


Date: *APRIL 21 79*



*MID-RANGE
OUTPUT
1.0 NSEC/
DIV*



*MAX
OUTPUT
1.0 NSEC/
DIV*



*MAX
OUTPUT
2.0 NSEC/DIV.*

- a) Output signal amplitude: *> 15 VOLTS (MAX)*
- b) Pulse width: *4 - 100 NSEC*
- c) Rise time: *≤ 100 PSEC*
- d) Fall time: *≤ 1.0 NSEC*
- e) PRF: *0 - 1 MHz FOR
MIN PW*
- f) Jitter, stability: *OK*
- g) Prime power: *+ 15 ± 1 VOLT
@ 210 mA*