


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AV14-MPT*

S.N.: *203*

Date: *APRIL 21 1979*



1.0 NSEC / DIV
≈ 8.0 VOLTS / DIV

- a) Output signal amplitude:
> 30 VOLTS
- b) Pulse width:
1.0 NSEC
(10% RISE TIME)
- c) Jitter, stability:
OK
- d) Prime power:
+15 ± 1 VOLT
185 mA