


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AVH-S*

S.N.: *206*

Date: *8/5/79*



a) Output signal amplitude:

*> 15 VOLTS*

b) Pulse width:

*0.5 nsec*

c) Jitter, stability:

*(10% RISE TIME)*

*OK*

d) Prime power:

*+15 ± 1 VOLT*

*150 mA*

*1.0 nsec/div*

*≈ 7 VOLTS/DIV.*