


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AUH-MPT*

S.N.: *208*

Date: *APRIL 26 79*



a) Output signal amplitude:

> 30 VOLTS

b) Pulse width:

≤ 1.0 NSEC

c) Jitter, stability:

(10% RISE TIME)

OK

d) Prime power:

+15 ± 1 VOLT

185 mA

2.0 NSEC/CM
≈ 8.0 VOLTS/CM