


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-S-N

S.N.: 215

Date: JUNE 26 79



$\approx 20 \text{ V/DIV}$   
 $1.0 \text{ NSEC/DIV}$

a) Output signal amplitude:

- 24 V

b) Pulse width:

$\leq 500 \text{ PSEC}$   
(10% RISE TIME)

c) Jitter, stability:

OK

d) Prime power:

$+15 \pm 1 \text{ VOLT}$

135 mA