



PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVI-U-A*
S.N.: *225*
Date: *JULY 27 '79*



*≈ 1.6 VOLTS/DIV
10 NS/CDIV*



*≈ 1.6 VOLTS/DIV
1.0 NS/CDIV*

- a) Output signal amplitude:
50 V (MAX)
- b) Pulse width:
0-100 NS/CDIV
- c) Rise time:
< 500 PSEC
- d) Fall time:
< 2.0 NS/CDIV
- e) PRF:
0-20 KHZ
- f) Jitter, stability:
6K
- g) Prime power:

*+15 ± 1.0 VOLT
220 MA*