


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AUP-AV-1-N*


S.N.: *231*

Date: *Aug 15 79*



*10 VOLTS/DIV
1.0 NSEC/DIV*

- a) Output signal amplitude:
-10 V (MAX)
- b) Pulse width:
0.2 → 4.0 NSEC
- c) Rise time:
< 100 PSEC
- d) Fall time:
< 100 PSEC
- e) PRF:
0 - 1.0 MPR
- f) Jitter, stability:
OK
- g) Prime power:



*10 VOLTS/DIV
1.0 NSEC/DIV*

*+15 ± 1 VOLT
275 mA*