


PULSE GENERATOR PERFORMANCE

CHECK


Model: *AVK-AV-1P*

S.N.: *232*

Date: *AUG 20 79*



*≈ 12 VOLTS/DIV (MAX OP)
1.0 NSEC/DIV*



*≈ 12 VOLTS/DIV
1.0 NSEC/DIV*

- a) Output signal amplitude:
+15 V (MAX)
- b) Pulse width:
0 - 50 NSEC
- c) Rise time:
≈ 100 PSEC
- d) Fall time: *(10% to 90%)*
≤ 1.0 NSEC
- e) PRF:
0 - 50 KHZ
- f) Jitter, stability:
OK
- g) Prime power:
*+15 ± 1 VOLT
275 mA*

