


PULSE GENERATOR

PERFORMANCE CHECK


Model: *AVK-AV-N-05*

S.N.: *239*

Date: *Aug 30 1979*



*1.0 NSEC/DIV
≈ 10 VOLTS/DIV*



*10 NSEC/DIV
≈ 10 VOLTS/DIV*

- a) Output signal amplitude:
0 TO -15 VOLTS
- b) Pulse width:
0 TO 50 NSEC
- c) Rise time:
*< 150 PSEC
(10% RISE)*
- d) Fall time:
< 2 NSEC
- e) PRF:
0 TO 50 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

*+15 ± 1 VOLT
280 mA.*