


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-V-A-N*

S.N.: *240*

Date: *MAR 30 1979*



5 NSEC / DIV
≈ 25 VOLTS / DIV

- a) Output signal amplitude:
0 TO -50 VOLTS
- b) Pulse width:
0 TO 100 NSEC
- c) Rise time:
< 0.5 NSEC
- d) Fall time:
< 2.0 NSEC
- e) PRF:
0 - 20 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

+15 ± 1 VOLT
240 mA