


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PVL-1AV-P*
S.N.: *244*
Date: *OCT 1 1979*



- a) Output signal amplitude:
> 250 VOLTS (MAX)
- b) Pulse width:
0 TO 100 NSEC
- c) Rise time:
≤ 5.0 NSEC
- d) Fall time:
≤ 5.0 NSEC
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:

10-0 NSEC/CM
≈ 125 VOLTS/DIV
(2.0 KHz)

+15 ± 1 VOLT
350 mA (5 KHz)