


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVM-MPT-P*

S.N.: *247*

Date: *OCT 15 1979*



*1.0 nsec / DIV
≈ 6.0 volts / DIV*

- a) Output signal amplitude:
> + 30 V
- b) Pulse width:
*≤ 1.0 nsec
(10% RISE TIME)*
- c) Rise time:
N/A
- d) Fall time:
N/A
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
*+15 ± 1.0 VOLT
195 mA*