


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVP-U-1*  
S.N.: *250*  
Date: *OCT 22 1979*



*1.0 nsec/div*  
*≈ 10.0 volts/div*  
*100 kHz*

- a) Output signal amplitude:  
*> 10 VOLTS*
- b) Pulse width:  
*0 TO 4.0 nsec*
- c) Rise time:  
*< 100 psec*
- d) Fall time:  
*< 100 psec*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*+15 ± 1 VOLT*

*300 mA*