



APPR  
UP 5/79

PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVM-1-P-10*

S.N.: *255*

Date: *NOV 16 1979*



- a) Output signal amplitude:  
*> 5.0 VOLTS (MAX)*
- b) Pulse width:  
*0 TO 6.0 NSEC.*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time: *(10%)*  
*≤ 100 PSEC*
- e) PRF:  
*0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*H: 1.0 nsec/cm*  
*V: ≈ 10 VOLTS/cm*  
*2.0 MHz*

↑  
*BEFORE  
UP 5/79*

*+24 VOLTS @*  
*410 mA (MAX)*